

Technical Bulletin

Acetaldehyde Assay Kit

Catalogue number MAK321

Product Description

Acetaldehyde is one of the most widely occurring aldehydes in nature and commonly used in industry. The metabolic byproduct of ethanol in the liver, acetaldehyde is toxic to the human body and rapidly converted to the less harmful acetic acid by the enzyme aldehyde dehydrogenase. People with a deficiency of aldehyde dehydrogenase accumulate acetaldehyde when consuming alcohol and this accumulation results in facial and body flushing often referred to as "Asian flush syndrome". Buildup of acetaldehyde has also been associated with the effects of hangovers from alcohol consumption. Although classified as a carcinogen, acetaldehyde is naturally found in many foods and beverages such as ripe fruit, coffee, and wine.

The acetaldehyde assay is based on the aldehyde dehydrogenase catalyzed oxidation of acetaldehyde, in which the generated NADH reduces a probe making it fluorescent. The fluorescence intensity of the product measured at $\lambda_{\text{ex}}=530~\text{nm}/\lambda_{\text{em}}=585~\text{nm}$ is directly proportional to acetaldehyde concentration in the sample.

The Acetaldehyde Assay Kit can be used to determine acetaldehyde in biological samples (e.g., plasma, serum, urine, tissue, and culture media.) or food/beverage samples (e.g., wine, coffee, and juice).

Components

The kit is sufficient for 100 fluorometric assays in 96-well plates.

•	Assay Buffer Catalog Number MAK321A	10 mL
•	NAD Solution Catalog Number MAK321B	1 mL
•	Probe Catalog Number MAK321C	750 µL
•	Enzyme A Catalog Number MAK321D	120 µL
•	Enzyme B Catalog Number MAK321E	120 µL
•	Acetaldehyde Standard, 3 M Catalog Number MAK321F	100 μL

Equipment Required but Not Provided

- Pipetting devices and accessories (for example, multichannel pipettor)
- Fluorescence plate reader
- Black flat-bottom 96 well plates
- 1.5 mL microcentrifuge tubes

Precautions and Disclaimer

For Research Use Only. Not for use in diagnostic procedures. Please consult the Safety Data Sheet for information regarding hazards and safe handling practices.

Storage/Stability

1

The kit is shipped on dry ice. Store components at -20 °C upon receipt. Standard may be stored at -20 °C to 4 °C.



Preparation Instructions

Reagent Preparation

Equilibrate Assay Buffer, NAD solution, and Probe to room temperature. Briefly centrifuge tubes before use. Keep Enzymes and Standard on ice.

Procedure

Use ultrapure water for the preparation of all Standards and Samples.

Acetaldehyde Standards

Prepare 1 mL of 15 mM Standard solution by mixing 5 μ L of the Acetaldehyde Standard (3 M) and 995 μ L of ultrapure water. Prepare 1 mL of 60 μ M Standard solution by mixing 4 μ L of 15 mM Standard solution with 996 μ L of ultrapure water. Prepare standards in 1.5 mL centrifuge tubes as described in Table 1.

Note: Prepare 60 μ M Standard solution and individual Standards fresh for each assay.

Table 1. Preparation of Acetaldehyde Standards

Tube	60 µM standard	Ultrapure water	Acetaldehyde (µM)
1	100 μL	0 μL	60
2	60 µL	40 μL	36
3	30 μL	70 μL	18
4	0 µL	100 μL	0

Transfer 50 μ L of each Standard into separate wells of a black, flat-bottom 96 well plate.

Sample Preparation

Clear and slightly colored Samples can be assayed directly. Biological fluid Samples (e.g., urine and serum) can be assayed directly after centrifuging to remove any particulates.

It is prudent to test several dilutions to determine an optimal dilution factor (n). Appropriate dilution in ultrapure water may be required.

Transfer 50 μ L of each sample in duplicate into separate wells (one well as "Sample" and one well as "Sample Blank").

Assay Reaction

 Set up the Master Reaction Mixes according to the scheme in Table 2. 50 µL of the appropriate Master Reaction Mix is required for all wells.

Table 2.Master Reaction Mixes

Reagent	Standards and Samples Volume	Sample Blank Volume
Assay Buffer	40 μL	41 µL
NAD Solution	8 µL	8 µL
Probe	5 μL	5 μL
Enzyme A	1 μL	_
Enzyme B	1 μL	1 μL

- 2. Add 50 μ L of the appropriate Master Reaction Mix to the Standards, the "Sample", and the "Sample Blank" wells.
- 3. Tap plate to mix briefly and thoroughly, then incubate for 30 minutes at room temperature. Read fluorescence at $\lambda_{ex} = 530 \text{ nm}/\lambda_{em} = 585 \text{ nm}$.

Results

Subtract the blank value (standard tube 4) from the values of the other standards. Plot the ΔF against standard concentrations and determine the slope of the standard curve. Calculate the acetaldehyde concentration of Sample(s):

Where:

 F_S = fluorescence reading of the Sample F_{SB} = fluorescence reading of the Sample Blank n = the sample dilution factor

Note: If the sample fluorescence value is higher than fluorescence for the 60 μ M acetaldehyde standard, dilute sample in ultrapure water and repeat the assay.

Conversions:

1 μ M acetaldehyde equals 44 μ g/L, or 44 ppb.

Notice

We provide information and advice to our customers on application technologies and regulatory matters to the best of our knowledge and ability, but without obligation or liability. Existing laws and regulations are to be observed in all cases by our customers. This also applies in respect to any rights of third parties. Our information and advice do not relieve our customers of their own responsibility for checking the suitability of our products for the envisaged purpose.

The information in this document is subject to change without notice and should not be construed as a commitment by the manufacturing or selling entity, or an affiliate. We assume no responsibility for any errors that may appear in this document.

Technical Assistance

Visit the tech service page at SigmaAldrich.com/techservice.

Terms and Conditions of Sale

Warranty, use restrictions, and other conditions of sale may be found at SigmaAldrich.com/terms.

Contact Information

For the location of the office nearest you, go to SigmaAldrich.com/offices.

The life science business of Merck operates as MilliporeSigma in the U.S. and Canada.

