



QUALITY CONTROL RANGES
Milliplex Rat Liver Injury Magnetic Panel
Kit Catalog# RLI1MAG-92K
Control Catalog # RLI1-6092 Lot# RL192-105 and RL192-205

Analyte	QC Level	Expected Range	Units	Analyte	QC Level	Expected Range	Units
ARG1	Control 1	0.80 – 1.7	ng/mL	5'-NT	Control 1	0.30 – 0.63	ng/mL
	Control 2	9.9 – 20	ng/mL		Control 2	3.8 – 7.9	ng/mL
GOT1	Control 1	0.39 – 1.2	ng/mL	SDH	Control 1	7.1 – 21	ng/mL
	Control 2	7.4 – 15	ng/mL		Control 2	61 – 127	ng/mL
GSTa	Control 1	0.14 – 0.33	ng/mL				
	Control 2	2.0 – 4.1	ng/mL				

RLI1-6092: RL192-105.205 (12 September 2024)

EMD MILLIPORE • 1-800-645-5476 • 1-800-645-5439 (FAX) www.emdmillipore.com

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA