

QUALITY CONTROL RANGES



MILLIPLEX® MAP Mouse CVD Magnetic Bead Panel 1
Kit Catalog # MCVD1MAG-77K
Control Catalog # MCVDMG1-6077-1
Lot# MCVD1-106 and MCVD1-206

<u>Analyte</u>	<u>QC Level</u>	<u>Expected Range</u>	<u>Units</u>
E-Selectin	Control I	1.2 - 2.4	ng/mL
	Control II	12 - 27	ng/mL
ICAM-1	Control I	0.18 - 0.38	ng/mL
	Control II	2.1 - 4.3	ng/mL
Pecam-1	Control I	0.54 - 1.1	ng/mL
	Control II	5.9 - 12	ng/mL
P-Selectin	Control I	11 - 24	ng/mL
	Control II	115 - 238	ng/mL
PAI-1Total	Control I	0.60 - 1.2	ng/mL
	Control II	6.1 - 13	ng/mL
proMMP-9	Control I	0.32 - 0.66	ng/mL
	Control II	3.4 - 7.1	ng/mL
Thrombomodulin	Control I	0.59 - 1.2	ng/mL
	Control II	6.5 - 14	ng/mL