



QUALITY CONTROL RANGES

PIIANP ELISA CATALOG # EZPIIANP-53K

STANDARD, LYOPHILIZED CATALOG#E8053-K

STANDARD	PIIANP VALUE
Lot Number: E53-16K	2341 ng/mL

The exact value of lyophilized standard is shown above.

Reconstitute with **0.2 mL** distilled or deionized water.

QUALITY CONTROLS 1 AND 2, LYOPHILIZED CATALOG #E6053-K

Reconstitute each vial with **0.2 mL** distilled or deionized water.

CONTROL	LOT NUMBER	EXPECTED RANGE	UNITS
QC 1	PIIA-116	69 - 144	ng/mL
QC 2	PIIA-216	415 - 861	ng/mL

E6053-K: PIIA-116.216 (12 April 2024)

EMD Millipore • 1-800-645-5476 • 1-800-645-5439 (FAX) • emdmillipore.com

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA