



QUALITY CONTROL RANGES
MILLIPLEX[®] MAP Human Fibrosis Panel 1
Catalog # HFIB1-100K
Control Catalog # HFIB1-6100

<u>Hormone</u>	<u>QC Level</u>	<u>Lot #</u>	<u>Expected Range</u>	<u>Units</u>
IL-18BPα	Control I	HFIB1-103	68 - 141	pg/mL
	Control II	HFIB1-203	333 - 692	pg/mL
COMP	Control I	HFIB1-103	249 - 517	pg/mL
	Control II	HFIB1-203	1268 - 2634	pg/mL
Uteroglobin	Control I	HFIB1-103	129 - 269	ng/mL
	Control II	HFIB1-203	641 - 1332	ng/mL

Note: The Quality Control Ranges are generated with overnight assay format using serum matrix provided in the kit.

Quality Control values in culture media are not tested.

HFIB1-6100: HFIB1-103.203 (26 June 2024)

EMD MILLIPORE • 1-800-645-5476 • 1-800-645-5439 (FAX) emdmillipore.com

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA