



QUALITY CONTROL RANGES

Mouse Aging Magnetic Bead Panel 1

Kit Catalog# MAGE1MAG-25K

Control Catalog# MAGE1-6025

Lot# MAGE1-103 and MAGE1-203

<u>Analyte</u>	<u>QC Level</u>	<u>Expected Range</u>	<u>Units</u>
CTACK	Control I	6.7 - 14	pg/mL
	Control II	79 - 165	pg/mL
FGF-21	Control I	95 - 197	pg/mL
	Control II	1064 - 2210	pg/mL
GnRH	Control I	32 - 66	pg/mL
	Control II	341 - 709	pg/mL
IL-6	Control I	97 - 201	pg/mL
	Control II	1054 - 2190	pg/mL
Jag1	Control I	3199 - 6645	pg/mL
	Control II	34823 - 72324	pg/mL
Leptin	Control I	45 - 94	pg/mL
	Control II	524 - 1088	pg/mL
Notch-1	Control I	1563 - 3247	ng/mL
	Control II	17550 - 36450	ng/mL
TNFa	Control I	30 - 63	pg/mL
	Control II	350 - 727	pg/mL

Note: The Quality Control Ranges are generated with overnight assay format using serum matrix provided in the kit. Control values in culture media are not tested.

MAGE1-6025: MAGE1-103.203 (12 November 2024)

EMD Millipore • 1-800-645-5476 • 1-800-645-5439 (FAX) • emdmillipore.com

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA