



QUALITY CONTROL RANGES

MILLIPLEX[®] MAP Human Amyloid Beta and Tau Magnetic Bead Panel
Kit Catalog # HNABTMAG-68K
Control Catalog # HABT-6068
Lot# HABT-110 and HABT-210

<u>Analyte</u>	<u>QC Level</u>	<u>Expected Range</u>	<u>Units</u>
A β 40	Control I	327 - 679	pg/mL
	Control II	1371 - 2848	pg/mL
A β 42	Control I	45 - 99	pg/mL
	Control II	166 - 345	pg/mL
tTau	Control I	163 - 339	pg/mL
	Control II	699 - 1451	pg/mL
pTau 181	Control I	12 - 24	pg/mL
	Control II	50 - 103	pg/mL

**Note: The Quality Control Ranges are generated with overnight assay format.
Quality Control values in culture media are not tested.**

HABT-6068: HABT-110.210 (17 JUNE 2025)

SigmaAldrich.com/support • SigmaAldrich.com

© 2025 Merck KGaA, Darmstadt, Germany and/or its affiliates. All Rights Reserved.

The vibrant M logo and MILLIPLEX are trademarks of Merck KGaA, Darmstadt Germany or its affiliates.