



QUALITY CONTROL RANGES

HUMAN ADIPONECTIN ELISA KIT CATALOG # EZHADP-61K, EZHADP-61BK

STANDARD, LYOPHILIZED CATALOG # E8061-K

STANDARD	HUMAN ADIPONECTIN VALUE
LOT NUMBER: HAE-21K	176 ng/mL

The exact value of lyophilized standard is shown above.

Reconstitute with **0.5 mL** distilled or deionized water.

QUALITY CONTROLS 1 AND 2, LYOPHILIZED CATALOG # E6361-K

Reconstitute each vial with **0.5 mL** distilled or deionized water.

CONTROL	LOT NUMBER	EXPECTED RANGE	UNITS
QC 1	HAE-121	2.4 – 5.0	ng/mL
QC 2	HAE-221	13 – 27	ng/mL

E6361-K: HAE-121.221 (15 August 2024)

EMD MILLIPORE • 1-800-645-5476 • 1-800-645-5439 (FAX) www.emdmillipore.com

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA