

Quick Start Guide

GenElute™ Blood Genomic DNA Miniprep Kit

NA2010, NA2020

Reagents to Prepare

- Dilute the Prewash Solution with 27.5 mL (70 prep) or 110 mL (350 prep) of 95-100% ethanol. After each use, tightly cap the diluted Prewash Solution to prevent ethanol evaporation.
- Dilute the Wash Solution with 80 mL (70 prep) or 360 mL (350 prep) of 95-100% ethanol. After each use, tightly cap the diluted Wash Solution to prevent ethanol evaporation.
- Dissolve Proteinase K in water to obtain a 20 mg/mL stock solution. This solution can be stored for several days at 2–8 °C. For longer-term storage, the unused portion of the solution may be stored in aliquots at –20 °C until needed.

Protocol

All spins at $\geq 12,000 \times g$.

Release DNA

1. Collect whole blood in anticoagulant tube (EDTA preferred).
2. Add 20 μ L of Proteinase K to 1.5 mL Microcentrifuge Tube and add up to 200 μ L whole blood sample. If sample is less than 200 μ L, dilute using Resuspension Solution. Vortex to thoroughly mix.
3. Add 200 μ L Lysis Solution C and vortex thoroughly to mix.
4. Incubate at 55° C for 10 min.
5. Add 200 μ L of 95-100% ethanol to lysate and vortex thoroughly to mix.

Prepare Column

6. Assemble Column with 2 mL collection tube.
7. Add 500 μ L Column Preparation Solution to column and spin for 1 min. Discard flowthrough.

Bind DNA to Column

8. Transfer the lysate into the prepared column. Spin 1 min. Discard collection tube with flowthrough and transfer the column to a fresh 2 mL collection tube..

Wash to Remove Contaminants

9. Add 500 μ L Prewash Solution to column. Spin 1 min. Discard flowthrough.
10. Add 500 μ L Wash Solution to column. Spin 3 min. Discard flowthrough.
11. Spin 1 min to dry column.

Elute Purified DNA

12. Transfer Column to new collection tube.
13. Add 200 μ L Elution Solution to center of column. Spin 1 min.
14. An additional elution may be performed if desired to improve yield.

Notice

We provide information and advice to our customers on application technologies and regulatory matters to the best of our knowledge and ability, but without obligation or liability. Existing laws and regulations are to be observed in all cases by our customers. This also applies in respect to any rights of third parties. Our information and advice do not relieve our customers of their own responsibility for checking the suitability of our products for the envisaged purpose.

The information in this document is subject to change without notice and should not be construed as a commitment by the manufacturing or selling entity, or an affiliate. We assume no responsibility for any errors that may appear in this document.

Contact Information

For the location of the office nearest you, go to SigmaAldrich.com/Offices.

Technical Assistance

Visit the tech service page on our web site at SigmaAldrich.com/TechService.

Standard Warranty

The applicable warranty for the products listed in this publication may be found at SigmaAldrich.com/Terms.