



QUALITY CONTROL RANGES

GLP-1 TOTAL ELISA CATALOG# EZGLP1T-36K, EZGLP1T-36BK

STANDARD, LYOPHILIZED CATALOG# E8036-K

STANDARD	GLP-1 TOTAL VALUE
Lot Number. 836-12K	1,000 pM

The exact value of lyophilized standard is shown above.

Reconstitute with **0.5 mL** distilled or deionized water.

QUALITY CONTROLS 1 AND 2, LYOPHILIZED CATALOG# E6036-K

Reconstitute each vial with **0.5 mL** distilled or deionized water.

CONTROL	LOT NUMBER	EXPECTED RANGE	UNITS
QC 1	GLP1-112	24 – 51	pM
QC 2	GLP1-212	143 – 298	pM

E6036-K: GLP1-112,212 (21 Jul 2020)

EMD Millipore • 1-800-645-5476 • 1-800-645-5439 (FAX) • emdmillipore.com

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA.