



QUALITY CONTROL RANGES

MILLIPLEX[®] MAP Human Immunoglobulin Isotyping Magnetic Bead Kit

Catalog # HGAMMAG-301K

Standard Catalog # HGAM-8301, Lot # 8301-14K

Control Catalog # PC-301

<u>Hormone</u>	<u>Lot #</u>	<u>Expected Range</u>	<u>Units</u>
IgM	301-14K	57 - 117	ng/mL
IgG1	301-14K	252 - 523	ng/mL
IgG2	301-14K	942 - 1957	ng/mL
IgG3	301-14K	3.7 - 7.6	ng/mL
IgG4	301-14K	3.1 - 14	ng/mL
IgA	301-14K	34 - 70	ng/mL

PC-301: 301-14K (rev. 27 March 2025)

EMD MILLIPORE • 1-800-645-5476 • 1-800-645-5439 (FAX) • www.emdmillipore.com

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA