Quick Start Guide

GenElute™ PCR Clean-Up Kit

NA1020

Reagents to Prepare

 Dilute the Wash Solution Concentrate with 48 mL of 100% ethanol. After each use, tightly cap the diluted Wash Solution to prevent ethanol evaporation.

Protocol

All spins at $12,000-16,000 \times g$.

Column Preparation

- 1. Assemble column (blue o-ring) with a collection tube.
- 2. Add 0.5 mL Column Preparation Solution to each column and spin for 30 sec to 1 min.

Sample Preparation

Add 5 volumes of Binding Solution to 1 volume of the PCR reaction and mix.

Bind DNA to Column

- 4. Transfer the solution into the column.
- 5. Spin for 1 min and discard the eluate.

Wash to Remove Contaminants

- Replace column with its collection tube and add 0.5 mL Wash Solution to the column and spin for 1 min. Discard the eluate.
- 7. Replace column with its collection tube and spin for an additional 2 min to remove excess ethanol. Discard the eluate and collection tube.

DNA Elution

- 8. Transfer column to fresh 2 mL collection tube.
- 9. Apply 50 μ L Elution Solution or water to the center of each column. Incubate at room temperature for 1 min.
- 10. Spin column for 1 min.

Notice

We provide information and advice to our customers on application technologies and regulatory matters to the best of our knowledge and ability, but without obligation or liability. Existing laws and regulations are to be observed in all cases by our customers. This also applies in respect to any rights of third parties. Our information and advice do not relieve our customers of their own responsibility for checking the suitability of our products for the envisaged purpose.

The information in this document is subject to change without notice and should not be construed as a commitment by the manufacturing or selling entity, or an affiliate. We assume no responsibility for any errors that may appear in this document.

Contact Information

For the location of the office nearest you, go to <u>SigmaAldrich.com/Offices</u>.

Technical Assistance

Visit the tech service page on our web site at SigmaAldrich.com/TechService.

Standard Warranty

The applicable warranty for the products listed in this publication may be found at <u>SigmaAldrich.com/Terms</u>.



