



QUALITY CONTROL RANGES
MILLIPLEX® MAP Human Soluble Cytokine Receptor Magnetic Bead Panel
Catalog # HSCRMAG-32K, HSCRMAG32KPX14

Control Catalog # HSCR-6032, Lot# HSCR-119.219

Hormone	QC Level	Expected Range	Units	Hormone	QC Level	Expected Range	Units
sCD30	Control 1	277 - 576	pg/mL	sIL-6R	Control 1	144 - 300	pg/mL
	Control 2	3041 - 6316	pg/mL		Control 2	1572 - 3265	pg/mL
sEGFR	Control 1	1489 - 3092	pg/mL	sRAGE	Control 1	152 - 316	pg/mL
	Control 2	15605 - 32410	pg/mL		Control 2	1612 - 3347	pg/mL
sgp130	Control 1	281 - 583	pg/mL	sTNFRI	Control 1	143 - 297	pg/mL
	Control 2	2961 - 6151	pg/mL		Control 2	1540 - 3198	pg/mL
sIL-1RI	Control 1	266 - 553	pg/mL	sTNFRII	Control 1	131 - 272	pg/mL
	Control 2	3007 - 6245	pg/mL		Control 2	1480 - 3074	pg/mL
sIL-1RII	Control 1	1185 - 2461	pg/mL	sVEGFR1	Control 1	1400 - 2908	pg/mL
	Control 2	13981 - 29038	pg/mL		Control 2	14459 - 30031	pg/mL
sIL-2R α	Control 1	269 - 558	pg/mL	sVEGFR2	Control 1	1373 - 2851	pg/mL
	Control 2	3030 - 6293	pg/mL		Control 2	15572 - 32343	pg/mL
sIL-4R	Control 1	142 - 294	pg/mL	sVEGFR3	Control 1	1418 - 2945	pg/mL
	Control 2	1591 - 3303	pg/mL		Control 2	14913 - 30974	pg/mL

Note: The Quality Control Ranges are generated with overnight assay format using serum matrix provided in the kit.

HSCR-6032: HSCR-119/219 (02 September 2025) • [SigmaAldrich.com/support](https://sigmaaldrich.com/support) • [SigmaAldrich.com](https://sigmaaldrich.com)