



QUALITY CONTROL RANGES
MILLIPLEX® MAP HIGF-I, II Magnetic Bead Panel
Kit Catalog # HIGFMAG-52K

Standard Catalog# HIGFMG-8052 Lot# HIGFM-12K

Analyte	Standard Value	Units
IGF-1	58,826	pg/mL
IGF-II	389,491	pg/mL

Control Catalog # HIGFMG-6052
Lot# HIGFM-112 and HIGFM-212

<u>Analyte</u>	<u>QC Level</u>	<u>Expected Range</u>	<u>Units</u>
IGF-I	Control I	746 - 1550	pg/mL
	Control II	3039 - 6312	pg/mL
IGF-II	Control I	4787 - 9941	pg/mL
	Control II	23719 - 49264	pg/mL

HIGFMG-6052: HIGFM-112.212 (12 MAY 2025) • SigmaAldrich.com/support • SigmaAldrich.com

© 2025 Merck KGaA, Darmstadt, Germany and/or its affiliates. All Rights Reserved. The vibrant M logo and MILLIPLEX are trademarks of Merck KGaA, Darmstadt Germany or its affiliates.