


Quick Start Guide

AAW™ 1-Channel Pipette

99900248 50 µL

99900246 1000 µL



AAW™ 1-Channel Pipette

The AAW™ 1-Channel Pipette is an electronic pipette designed for the AAW™ Automated Assay Workstation (AAW™ Workstation) liquid handling robot.

User Guide

Read the AAW™ Workstation user guide before attempting to operate. It can be downloaded from the product page at SigmaAldrich.com. The user guide contains detailed information about this product, modules, and accessories.

Pipette Capacity

AAW™ 1-Channel Pipettes are available in 50 µL and 1000 µL sizes. The minimum/maximum volumetric capacities of each type are listed below.


- 50 µL pipette: 1–50 µL
- 1000 µL pipette: 5–1000 µL

Tip Compatibility

AAW™ 1-Channel Pipettes are optimized for use with Flex 50 µL, 200 µL, and 1000 µL pipette tips. AAW™ Workstation pipettes are not recommended for use with tips from other manufacturers. For more information, see the System Description and Labware chapters in the AAW™ Workstation User Guide.

Purchase Flex pipette tips at SigmaAldrich.com.

Box Contents



Installation

For assistance with installation, contact SigmaAldrich.com/TechService.

Attach

Instructions on the touchscreen, or in the Opentrons App, will guide you through the attachment process. On the touchscreen, tap **Instruments** and then tap an empty mount. To attach your pipette:

1. Clear the deck area so it is free from obstructions. The gantry will move during the installation process.
2. Connect the pipette to the left or right mount on the gantry. To attach, align the mounting holes on the back of the pipette with the pins on the gantry.
3. Secure the pipette by tightening the two 2.5 mm mounting screws. These are captive screws recessed into holes located on the front of the pipette.

Calibrate

The pipette requires calibration before use. Instructions on the touchscreen, or in the Opentrons App, will guide you through the calibration process.

To calibrate your pipette:

1. Clear the deck area so it is free from obstructions. The gantry and pipette will move during calibration.
2. Attach the calibration probe to the pipette and start the automated calibration procedure.
3. Remove and store the probe when calibration is complete. Your pipette is ready for use.

Clean

Clean the outside of the pipette with a soft cloth and isopropyl alcohol or ethyl alcohol. Do not use acetone or other aggressive solvents. For more cleaning and decontamination information, see the Pipette Cleaning section in the AAW™ Workstation User Guide.

Notice

We provide information and advice to our customers on application technologies and regulatory matters to the best of our knowledge and ability, but without obligation or liability. Existing laws and regulations are to be observed in all cases by our customers. This also applies in respect to any rights of third parties. Our information and advice do not relieve our customers of their own responsibility for checking the suitability of our products for the envisaged purpose.

The information in this document is subject to change without notice and should not be construed as a commitment by the manufacturing or selling entity, or an affiliate. We assume no responsibility for any errors that may appear in this document.

Technical Assistance

Visit the tech service page on our web site at SigmaAldrich.com/TechService.

Terms and Conditions of Sale

Warranty, use restrictions, and other conditions of sale may be found at SigmaAldrich.com/Terms.

Contact Information

For the location of the office nearest you, go to SigmaAldrich.com/Offices.

MilliporeSigma, Millipore, AAW, and Sigma-Aldrich are trademarks of Merck KGaA, Darmstadt, Germany or its affiliates. All other trademarks are the property of their respective owners. Detailed information on trademarks is available via publicly accessible resources.
© 2025 Merck KGaA, Darmstadt, Germany and/or its affiliates. All Rights Reserved.

The life science business of Merck KGaA, Darmstadt, Germany operates as MilliporeSigma in the U.S. and Canada.

