



QUALITY CONTROL RANGES

Human IL-18 Single plex Magnetic Bead Panel
Kit Catalog#: HIL18MAG-66K

Control Catalog#: HIL18-6066
Lot#: H1866-106 and H1866-206

<u>Analyte</u>	<u>QC Level</u>	<u>Expected Range</u>	<u>Units</u>
IL-18	Control 1	250 – 520	pg/mL
	Control 2	959 – 1991	pg/mL

HIL18-6066: H1866-106.206 (03FEB2025)

EMD MILLIPORE • 1-800-645-5476 • 1-800-645-5439 (FAX) www.emdmillipore.com

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA