



Quality Control Ranges
Human Neuroscience Magnetic Bead Panel 2
Kit Cat# HNS2MAG-95K

Control Cat# HNS2-6095-2
Lot# HNS2-106 and HNS2-206

<u>Analyte</u>	<u>Control Level</u>	<u>Expected Range</u>	<u>Units</u>
TREM2	Control 1	333 - 691	pg/mL
	Control 2	1721 - 3575	pg/mL
ApoE4	Control 1	6023 - 12508	pg/mL
	Control 2	30282 - 62893	pg/mL
Angiogenin	Control 1	61 - 127	pg/mL
	Control 2	325 - 675	pg/mL
FABP3	Control 1	584 - 1213	pg/mL
	Control 2	3270 - 6791	pg/mL
Neurogranin	Control 1	76 - 157	pg/mL
	Control 2	480 - 996	pg/mL
Ferritin	Control 1	156 - 323	pg/mL
	Control 2	762 - 1583	pg/mL

HNS2-6095-2: HNS2-106.206 (12 March 2024)

EMD Millipore • 1-800-645-5476 • 1-800-645-5439 (FAX) • emdmillipore.com

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA