



Product Information

SIGMA QUALITY CONTROL TEST PROCEDURE

Enzymatic Assay of PHOSPHOGLUCOSE ISOMERASE¹ (EC 5.3.1.9)

PRINCIPLE:

D-Fructose 6-Phosphate $\xrightarrow{\text{PGI}}$ D-Glucose 6-Phosphate

D-Glucose 6-Phosphate + β -NADP $\xrightarrow{\text{G-6-PDH}}$ 6-Phosphogluconate + β -NADPH

Abbreviations used:

PGI = Phosphoglucose Isomerase

β -NADPH = β -Nicotinamide Adenine Dinucleotide Phosphate, Reduced Form

β -NADP = β -Nicotinamide Adenine Dinucleotide Phosphate, Oxidized Form

G-6-PDH = Glucose-6-Phosphate Dehydrogenase

CONDITIONS: T = 25°C, pH = 7.4, A_{340nm}, Light path = 1 cm

METHOD: Continuous Spectrophotometric Rate Determination

REAGENTS:

- A. 250 mM Glycylglycine Buffer, pH 7.4 at 25°
(Prepare 100 ml in deionized water using Glycylglycine, Free Base, Sigma Prod. No. G-1002. Adjust the pH to 7.4 with 1 M NaOH.)
- B. 100 mM D-Fructose 6-Phosphate Solution (F-6-P)
(Prepare 1 ml in deionized water using D-Fructose 6-Phosphate, Disodium, Sigma Prod. No. F-3627.)
- C. 20 mM β -Nicotinamide Adenine Dinucleotide Phosphate Solution (NADP)
(Prepare 1 ml in deionized water using β -Nicotinamide Adenine Dinucleotide Phosphate, Sodium Salt, Sigma Prod. No. N-0505. **PREPARE FRESH.**)
- D. 100 mM Magnesium Chloride Solution (MgCl₂)
(Prepare 10 ml in deionized water using Magnesium Chloride, Hexahydrate, Sigma Prod. No. M-0250.)

Enzymatic Assay of PHOSPHOGLUCOSE ISOMERASE¹ (EC 5.3.1.9)

REAGENTS: (continued)

- E. Glucose-6-Phosphate Dehydrogenase Enzyme Solution (G-6-PDH)
(Immediately before use, prepare a solution containing 50 units/ml of Glucose-6-Phosphate Dehydrogenase, Sigma Prod. No. G-6378, in cold deionized water.)
- F. Phosphoglucose Isomerase Enzyme Solution (PGI)
(Immediately before use, prepare a solution containing 0.3 - 0.7 unit/ml in cold deionized water.)

PROCEDURE:

Pipette (in milliliters) the following reagents into suitable cuvettes:

	<u>Test</u>	<u>Blank</u>
Deionized Water	2.00	2.00
Reagent A (Buffer)	0.50	0.50
Reagent B (F-6-P)	0.10	0.10
Reagent C (NADP)	0.10	0.10
Reagent D (MgCl ₂)	0.10	0.10
Reagent E (G-6-PDH)	0.10	0.10

Mix by inversion and equilibrate to 25°C. Monitor the $A_{340\text{nm}}$ until constant, using a suitably thermostatted spectrophotometer. Then add:

Reagent F (PGI)	0.10	-----
Deionized Water	-----	0.10

Immediately mix by inversion and record the increase in $A_{340\text{nm}}$ for approximately 5 minutes. Obtain the $\Delta A_{340\text{nm}}/\text{minute}$ using the maximum linear rate for both the Test and Blank.

CALCULATIONS:

$$\text{Units/ml enzyme} = \frac{(\Delta A_{340\text{nm}}/\text{min Test} - \Delta A_{340\text{nm}}/\text{min Blank})(3)(\text{df})}{(6.22)(0.1)}$$

3 = Total volume (in milliliters) of assay

df = Dilution factor

6.22 = Millimolar extinction coefficient of β -NADPH at 340 nm

0.1 = Volume (in milliliter) of enzyme used

Enzymatic Assay of PHOSPHOGLUCOSE ISOMERASE¹ (EC 5.3.1.9)

CALCULATIONS: (continued)

$$\text{Units/mg solid} = \frac{\text{units/ml enzyme}}{\text{mg solid/ml enzyme}}$$

$$\text{Units/mg protein} = \frac{\text{units/ml enzyme}}{\text{mg protein/ml enzyme}}$$

UNIT DEFINITION:

One unit will convert 1.0 μ mole of D-fructose 6-phosphate to D-glucose 6-phosphate per minute at pH 7.4 at 25°C.

FINAL ASSAY CONCENTRATION:

In a 3.00 ml reaction mix, the final concentrations are 42 mM glycylglycine, 3.3 mM D-fructose 6-phosphate, 0.67 mM β -nicotinamide adenine dinucleotide phosphate, 3.3 mM MgCl_2 , 5.0 units glucose-6-phosphate dehydrogenase and 0.03 - 0.07 unit phosphoglucose isomerase.

REFERENCE:

Bergmeyer, H.U., Gawehn, K., and Grassl, M. (1974) in *Methods of Enzymatic Analysis* (Bergmeyer, H.U., ed.) Volume 1, 2nd ed., 501-503, Academic Press, Inc., New York, NY

NOTES:

1. This assay procedure is not to be used to assay the activities of Phosphoglucose Isomerase from *Bacillus stearothermophilus*, Sigma Prod. No. P-5538.
2. Glucose-6-Phosphate Dehydrogenase unit definition: One unit will oxidize 1.0 μ mole of D-glucose 6-phosphate to 6-phospho-D-gluconate per minute in the presence of NADP at pH 7.4 at 25°C.
3. This assay is based on the cited reference.
4. Where Sigma Product or Stock numbers are specified, equivalent reagents may be substituted.

Sigma warrants that the above procedure information is currently utilized at Sigma and that Sigma products conform to the information in Sigma publications. Purchaser must determine the suitability of the information and products for its particular use. Upon purchase of Sigma products, see reverse side of invoice or packing slip for additional terms and conditions of sale.