

QUALITY CONTROL RANGES



MILLIPLEX® MAP Mouse Acute Phase Magnetic Bead Panel 2 Catalog # MAP2MAG-76K

Standard Catalog # MAP2-8076-2 Lot# 8762-7K

Analytes	Standard Value
Adipsin	2.34 ng/mL
AGP	0.378 ng/mL
A2M	5223 ng/mL

Analytes	Standard Value
CRP	15.5 ng/mL
Haptoglobin	225 ng/mL
SAP	2278 ng/mL

Control Catalog # MAP2-6076-2

<u>Apolipoproteins</u>	<u>QC Level</u>	<u>Lot #</u>	<u>Expected Range</u>	<u>Units</u>
Adipsin	Control I	MAP2-107	0.021 – 0.044	ng/mL
	Control II	MAP2-207	0.13 – 0.27	ng/mL
AGP	Control I	MAP2-107	0.004 – 0.007	ng/mL
	Control II	MAP2-207	0.024 – 0.049	ng/mL
A2M	Control I	MAP2-107	52 – 108	ng/mL
	Control II	MAP2-207	316 – 657	ng/mL
CRP	Control I	MAP2-107	0.14 – 0.30	ng/mL
	Control II	MAP2-207	0.92 – 1.9	ng/mL
Haptoglobin	Control I	MAP2-107	2.1 – 4.4	ng/mL
	Control II	MAP2-207	14 – 29	ng/mL
SAP	Control I	MAP2-107	18 – 38	ng/mL
	Control II	MAP2-207	123 – 255	ng/mL

MAP2-6076-2: MAP2-107.207 (22 May 2023)

EMD MILLIPORE • 1-800-645-5476 • 1-800-645-5439 (FAX) www.emdmillipore.com

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA