

Quality Control Ranges

MILLIPLEX® Human Neurological Disorders Magnetic Bead Panel 3

Catalog Number **HND3MAG-39K**

QC Catalog Number/ Lot Number **HND3-6039-3/ HND3-108 and HND3-208**

<u>Analyte</u>	<u>QC Level</u>	<u>Expected Range</u>	<u>Units</u>
SOD1	Control I	0.32 - 0.66	ng/mL
	Control II	3.2 - 6.6	ng/mL
Angiotensinogen	Control I	9.6 - 20	ng/mL
	Control II	103 - 215	ng/mL
Fetuin A	Control I	30 - 63	ng/mL
	Control II	315 - 654	ng/mL
SOD2	Control I	0.24 - 0.51	ng/mL
	Control II	2.5 - 5.1	ng/mL
Contactin-1	Control I	0.60 - 1.3	ng/mL
	Control II	6.5 - 13	ng/mL
Kallikrein-6	Control I	0.29 - 0.60	ng/mL
	Control II	3.1 - 6.5	ng/mL
Osteopontin	Control I	1.0 - 2.2	ng/mL
	Control II	11 - 22	ng/mL

The life science business of Merck KGaA, Darmstadt, Germany operates as MilliporeSigma in the U.S. and Canada.

The vibrant M, Milliplex, Millipore and Sigma-Aldrich are trademarks of Merck KGaA, Darmstadt, Germany or its affiliates.

© 2026 Merck KGaA, Darmstadt, Germany and/or its affiliates. All Rights Reserved.

Cat#: HND3-6039-3, Lot#: HND3-108 and HND3-208 (06 MAR 2026)

