



QUALITY CONTROL RANGES
MILLIPLEX® MAP Human CVD Panel 3 (Acute Phase) Magnetic Bead Panel
Catalog # HCVD3MAG-67K
Control Catalog # HCVMG3-6067-3 Lot # HCVM3-112 and HCVM3-212

<u>Hormone</u>	<u>QC Level</u>	<u>Expected Range</u>	<u>Units</u>
A2M	Control 1	6.1 - 13	ng/mL
	Control 2	76 - 204	ng/mL
CRP	Control 1	0.12 - 0.26	ng/mL
	Control 2	1.4 - 2.9	ng/mL
Fetuin A	Control 1	3.0 - 6.3	ng/mL
	Control 2	32 - 65	ng/mL
AGP	Control 1	1.2 - 2.4	ng/mL
	Control 2	13 - 26	ng/mL
Fibrinogen	Control 1	0.31 - 0.64	ng/mL
	Control 2	3.2 - 6.6	ng/mL
L-Selectin	Control 1	0.05 - 0.09	ng/mL
	Control 2	0.47 - 1.0	ng/mL
SAP	Control 1	0.04 - 0.09	ng/mL
	Control 2	0.45 - 0.9	ng/mL
Haptoglobin	Control 1	0.73 - 1.5	ng/mL
	Control 2	7.9 - 16	ng/mL
PF4/CXCL4	Control 1	0.62 - 1.3	ng/mL
	Control 2	6.3 - 13	ng/mL
Adipsin	Control 1	1.0 - 2.2	ng/mL
	Control 2	13 - 26	ng/mL

Note: The Quality Control Ranges are generated with overnight assay format.