

QUALITY CONTROL RANGES



MILLIPLEX[®] MAP Human TIMP Panel 1 Magnetic Kit Catalog # HTMP1MAG-54K

MILLIPLEX[®] MAP Human TIMP Panel 2 Magnetic Kit Catalog # HTMP2MAG-54K

Control Catalog # HTIMP-6054 Lot# HTMP-111A and HTMP-211A

<u>Hormone</u>	<u>QC Level</u>	<u>Expected Range</u>	<u>Units</u>
HTIMP-1	Control I	263 - 546	pg/mL
	Control II	1260 - 2617	pg/mL
HTIMP-2	Control I	659 - 1368	pg/mL
	Control II	3246 - 6741	pg/mL
HTIMP-3	Control I	1484 - 3082	pg/mL
	Control II	6865 - 14257	pg/mL
HTIMP-4	Control I	154 - 320	pg/mL
	Control II	768 - 1596	pg/mL

Note: The Quality Control Ranges are generated with overnight assay format using assay buffer provided in the kit. Quality Control values in culture media are not tested.

HTIMP-6054; HTMP-111A.211A (27 July 2021)

EMD MILLIPORE • 1-800-645-5476 • 1-800-645-5439 (FAX) • www.emdmillipore.com

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA.